

Summary of an evaluation provided by SKUP | Flowflex SARS-CoV-2 Ag Rapid Test

Manufacturer Acon Biotech Hangzhou Co. Ltd.

Supplier Acon Biotech Hangzhou Co. Ltd.
(requesting company)

Launched in Scandinavia 2020



Aim

To assess the diagnostic performance and user-friendliness of Flowflex SARS-CoV-2 Ag Rapid Test when used under real life conditions by intended users in dedicated COVID-19 test centres.

Examination

Recommended goals and results

Overall diagnostic sensitivity

WHO recommends a minimum performance requirement of $\geq 80\%$ sensitivity compared to a nucleic acid-amplification test (NAAT) reference assay

Overall diagnostic sensitivity was not met: 75 % (90 % CI: 68-82 %)*

Overall diagnostic specificity

WHO recommends a minimum performance requirement of $\geq 97\%$ specificity compared to a NAAT reference assay

Overall diagnostic specificity was met: 99,6 % (90 % CI: 98,6-99,9 %)*

User-friendliness

Quality goal; a total rating of 'Satisfactory' by SKUP

The quality goal of user-friendliness was fulfilled

Background

Measurement system

In vitro diagnostic rapid test for qualitative detection of SARS-CoV-2

Intended users

Health care professionals

Sample material

Nasal or nasopharyngeal specimen, of which the first was evaluated by SKUP

Material and methods

Participants

564 persons exposed to individuals with confirmed SARS-CoV-2 infection, of whom 121 (21 %) tested positive on one of the comparison methods.

Comparison method

A real time polymerase chain reaction (RT-PCR) method for detection of SARS-CoV-2 at the Clinical Diagnostic Department at the Hospital of South West Jutland in Esbjerg and the Department of Clinical Biochemistry at Bispebjerg Hospital in Copenhagen NV.

Analytical procedure

Subjects exposed to an individual with confirmed SARS-CoV-2 infection were invited to participate in the evaluation. The sampling procedure, performed by trained health care professionals, included one oropharyngeal swab sample for RT-PCR detection and one nasal swab sample from both nostrils, for the Flowflex SARS-CoV-2 Ag Rapid Test. The oropharyngeal swab for RT-PCR detection was immediately placed into sterile tubes, containing 2-3 mL of viral transport media, until transported to the clinical laboratory.

The nasal swab was placed into the test vial containing extraction buffer and analysed in accordance with the instructions from the manufacturer. Three lots of Flowflex SARS-CoV-2 Ag Rapid Test were used.

User-friendliness

Assessed by trained health care professionals using a questionnaire with three given ratings; satisfactory, intermediate and unsatisfactory.

Additional results

Sensitivity stratified on cycle threshold (ct) values for the E-gene:

<33: 78 %: (90 % CI: 70-84 %)*

<30: 82 %: (90 % CI: 75-88 %)*

<25: 83 %: (90 % CI: 76-89 %)*

Prevalence:

21 %

Positive predictive value (PPV):

95 %

Negative predictive value (NPV):

97 %

Acon Biotech Hangzhou Co. Ltd. has accepted the report without further comments

**Confidence interval (CI) is for information only*